

# Declaration of Conformity

## MANUFACTURER'S DECLARATION OF CONFORMITY

AUSTRALIAN THERAPEUTIC GOODS (MEDICAL DEVICES) REGULATIONS 2002

### FULL QUALITY ASSURANCE PROCEDURES


This is a declaration of conformity made under clause 1.8 of Schedule 3 to the Therapeutic Goods (Medical Devices) Regulations 2002.

**Manufacturer's name:** SunTech Medical, Inc.  
**Business address:** 5827 South Miami Blvd, Suite 100, Morrisville, NC 27560  
**Medical device(s):** Blood Pressure Hoses  
**Classification:** I (non-measuring, non-sterile)  
**GMDN code and term:** 61228 - External cuff/pad connection tubing  
**Scope of application:** 91-0028-XX (where -XX indicates any number 00 to 99)  
**Basic UDI:** 084093510000000000H100BT

Each kind of medical device to which the system has been applied complies with the applicable provisions of the essential principles, the classification rules, and the full quality assurance procedures, at each stage, from the design of the device until its final inspection before being supplied.

**Full quality assurance procedures certificate:** See attached certificate  
**Design examination certificate (if applicable):** None  
**Standards applied:** See attached Schedule A

#### Authorised signatory:

DocuSigned by:  
*Tonia Bryant*  
 Signer Name: Tonia Bryant  
 Signing Reason: I approve this document  
 Signing Time: 6/20/2022 | 8:47:11 AM PDT  
 74D91508594B47A18C3113C71002CECD

6/20/2022

Signature  
 Tonia E. Bryant, Regulatory Affairs Manager

Date

**Declaration of Conformity - Schedule A**

**Standards Applied:**

Cleaning/Disinfection	ISO 17664:2017	Processing of health care products - Information to be provided by the medical device manufacturer for the processing of medical devices
Biocompatibility	EN ISO 10993-1:2018	Biological Evaluation of Medical Devices-Part 1: Evaluation and testing within a risk management process
	EN ISO 10993-5:2009	Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity
	ISO 10993-10:2010	Biological evaluation of medical devices-Part 10: Tests for irritation and skin sensitization
	ISO 10993-23:2021	Biological evaluation of medical devices - Part 23: Tests for irritation - First Edition
Symbols	ISO 15223-1:2021	Medical Devices - Symbols to be used with information to be supplied by the manufacturer - Part 1: General requirements
Information	ISO 20417:2021	Medical devices - Information to be supplied by the manufacturer
Sphygmomanometers	IEC 80601-2-30:2018	Medical Electrical Equipment - Part 2-30: Particular Requirements for the Basic Safety and Essential Performance of Automated Non-Invasive Sphygmomanometers
Quality System	EN ISO 13485:2016	Medical Devices - Quality Management Systems - Requirements for Regulatory Purposes
Risk Management	EN ISO 14971:2019	Medical devices - Application of risk management to medical devices