

# Declaration of Conformity

## MANUFACTURER'S DECLARATION OF CONFORMITY

*AUSTRALIAN THERAPEUTIC GOODS (MEDICAL DEVICES) REGULATIONS 2002*

### FULL QUALITY ASSURANCE PROCEDURES

This is a declaration of conformity made under clause 6.6 of Schedule 3 to the Therapeutic Goods (Medical Devices) Regulations 2002.


**Manufacturer's name:** SunTech Medical, Inc.  
**Business address:** 5827 South Miami Blvd, Suite 100, Morrisville, NC 27560  
**Medical device(s):** Eclipse (D-Ring) Cuff  
**Classification:** I (non-measuring, non-sterile)  
**GMDN code and term:** 34978 – Blood pressure cuff, reusable  
**Scope of application:** 98-0068-XX (where -XX indicates any number 00 to 99)  
**Basic UDI:** 0840935100000000SELF100EX

Each kind of medical device to which the technical documentation applies complies with the applicable provisions of the essential principles and the classification rules before being supplied.

**Full quality assurance procedures certificate:** *ISO 13485:2016*  
*Intertek Testing Services NA, Inc. – Certificate No. 0068171-04*

**Standards applied:** See attached Schedule A

#### Authorised signatory:

Signed by:  
*Tonia Bryant*  
  
 Signer Name: Tonia Bryant  
 Signing Reason: I approve this document  
 Signing Time: 9/3/2024 | 10:26:38 AM PDT  
 Signature  
 Tonia E. Bryant, Regulatory Affairs Director  
 74D91508594B47A18C3113C71002CECD

9/3/2024

Date

## Declaration of Conformity – Schedule A

### Standards Applied:

Cleaning/Disinfection	ISO 17664:2017	Processing of health care products – Information to be provided by the medical device manufacturer for the processing of medical devices
Safety	IEC 80601-2-30:2009 + A1:2013	Medical Electrical Equipment - Part 2-30: Particular Requirements For The Basic Safety And Essential Performance of sphygmomanometers
	EN1060-3: 1997 + A2: 2009	Non-invasive sphygmomanometers-Part 3: Supplementary requirements for electro-mechanical blood pressure measuring systems
	ISO 80369-5:2016	Small-bore connectors for liquids and gases in healthcare applications - Part 5: Connectors for limb cuff inflation applications
Biocompatibility	EN ISO 10993-1:2018	Biological Evaluation of Medical Devices-Part 1: Evaluation and testing within a risk management process
	EN ISO 10993-5:2009	Biological evaluation of medical devices – Part 5: Tests for in vitro cytotoxicity
	ISO 10993-10:2010	Biological evaluation of medical devices-Part 10: Tests for irritation and skin sensitization
Symbols	ISO 15223-1:2021	Medical Devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements
Information	ISO 20417:2021	Medical devices – Information to be supplied by the manufacturer
Quality System	EN ISO 13485:2016	Medical Devices – Quality Management Systems – Requirements for Regulatory Purposes
Risk Management	EN ISO 14971:2019	Medical devices – Application of risk management to medical devices