

Declaration of Conformity

MANUFACTURER'S DECLARATION OF CONFORMITY

AUSTRALIAN THERAPEUTIC GOODS (MEDICAL DEVICES) REGULATIONS 2002

FULL QUALITY ASSURANCE PROCEDURES

This is a declaration of conformity made under clause 1.8 of Schedule 3 to the Therapeutic Goods (Medical Devices) Regulations 2002.

Manufacturer's name: SunTech Medical, Inc.

Business address: 5827 South Miami Blvd, Suite 100, Morrisville, NC 27560

Medical device(s): Non-Invasive Oscillmoetric Spot Check Vital Signs device with optional Temperature and Pulse

Oximetry, CT40, Model 260

Classification: IIa

GMDN code and term: 57960 – Multiple physiological parameters spot-check system, clinical 45607 (SpO2) – Pulse

oximeter

Scope of application: 99-0134-XX (where -XX indicates any number 00 to 99)

Basic UDI: 084093510000000000002607H

Each kind of medical device to which the system has been applied complies with the applicable provisions of the essential principles, the classification rules, and the full quality assurance procedures, at each stage, from the design of the device until its final inspection before being supplied.

Full quality assurance procedures

certificate:

ISO 13485:2016 / MDSAP Certificate 0068171-05

Design examination certificate

(if applicable):

None

Standards applied: See attached Schedule A

Authorised signatory:

—Signed by: Tonia Bryant

U

Signer Name: Tonia Bryant

Signing Reason: I approve this document Signing Time: 3/24/2025 | 11:53:22 AM PDT

-74D91508594B47A18C3113C71002CECD

Tonia E. Bryant, Regulatory Affairs Dire

Date

3/24/2025

Declaration of Conformity - Schedule A

99-0131-XX-RA7-EN-RevB Page 1 of 2



Standards Applied:

| Safety | IEC 60601-1: Ed. 3.1 (2020) | Medical electrical equipment – Part 1: General requirements for basic safety and essential performance |
|------------------------------|--|---|
| Performance/Safety | IEC80601-2-30: Ed. 2.0 (2018) | Medical electrical equipment – Part 2-30: Particular requirements for the basic safety and essential performance of automated non-invasive sphygmomanometers |
| | ISO 80601-2- 56:2017 | Medical electrical equipment — Part 2-56: Particular requirements for basic safety and essential performance of clinical thermometers for body temperature measurement |
| | ISO 80601-2-61: Ed.2.0 (2017) | Medical electrical equipment — Part 2-61: Particular requirements for basic safety and essential performance of pulse oximeter equipment |
| | ISO 81060-1: 2012 | Non-invasive sphygmomanometers — Part 1: Requirements and test methods for non-automated measurement |
| EMC/EMI/ ESD | IEC 60601-1-2: Ed. 4.0 (2020) | Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests |
| Software | IEC 62304: Ed. 1.1 (2015) | Medical device software – Software life cycle processes |
| Usability | IEC 60601-1-6:2010 +A MD1:2013+AMD2: 2020 | Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability |
| | EN 62366-1: 2015 | Medical devices – Part 1: Application of usability engineering to medical devices |
| Clinical | IEC 81060-2: 2013 | Non-Invasive sphygmomanometers - Part 2 Clinical investigation of intermittent automated measurement type |
| Biocompatibility (System) | ISO 10993-1: 2018 | Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process |
| Risk Management | ISO 14971:2019 | Medical devices — Application of risk management to medical devices |
| Quality System | ISO 13485: 2016 | Medical devices – Quality management systems – Requirements for regulatory purposes |
| Symbols | ISO 15223-1:2021 | Medical Devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements |
| Information | ISO 20417:2021 | Medical devices - Information to be supplied by the manufacturer |

99-0131-XX-RA7-EN-RevB Page 2 of 2