

# Declaration of Conformity

## MANUFACTURER'S DECLARATION OF CONFORMITY

AUSTRALIAN THERAPEUTIC GOODS (MEDICAL DEVICES) REGULATIONS 2002

### FULL QUALITY ASSURANCE PROCEDURES

This is a declaration of conformity made under clause 1.8 of Schedule 3 to the Therapeutic Goods (Medical Devices) Regulations 2002.

**Manufacturer's name:** SunTech Medical, Inc.

**Business address:** 5827 South Miami Blvd, Suite 100, Morrisville, NC 27560

**Medical device(s):** Non-Invasive Ambulatory Blood Pressure device, Oscar 2, Model 250D (Bravo Mini)

**Classification:** IIa

**GMDN code and term:** 36888 – Blood pressure ambulatory recorder

**Scope of application:** 99-0233-XX (where -XX indicates any number 00 to 99); NJXXX (where -XXX indicates any number 000 to 999)

**Basic UDI:** 084093510000000000250D92

Each kind of medical device to which the system has been applied complies with the applicable provisions of the essential principles, the classification rules, and the full quality assurance procedures, at each stage, from the design of the device until its final inspection before being supplied.


**Full quality assurance procedures certificate:** ISO 13485:2016 / MDSAP Certificate 0068171-05

**Design examination certificate (if applicable):** None

**Standards applied:** See attached Schedule A

#### Authorised signatory:

Signed by:  
*Tonia Bryant*

 Signer Name: Tonia Bryant  
 Signing Reason: I approve this document  
 Signing Time: 3/24/2025 | 11:53:55 AM PDT  
 74D91508594B47A18C3113C71002CECD

3/24/2025

Tonia E. Bryant, Regulatory Affairs Director

Date

#### Declaration of Conformity – Schedule A

#### Standards Applied:

Safety	IEC 60601-1: Ed. 3.2 (2020)	Medical electrical equipment – Part 1: General requirements for basic safety and essential performance
Performance/Safety	IEC80601-2-30: Ed. 2.0 (2018)	Medical electrical equipment – Part 2-30: Particular requirements for the basic safety and essential performance of automated non-invasive sphygmomanometers
	ISO 81060-1: 2012	Non-invasive sphygmomanometers — Part 1: Requirements and test methods for non-automated measurement
	IEC 60601-1-11:2020	Medical electrical equipment - Part 1-11: General requirements for basic safety and essential performance — Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment
EMC/EMI/ESD	IEC 60601-1-2: 2020	Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral Standard: Electromagnetic disturbances – Requirements and tests
Software	IEC 62304: Ed. 1.1 (2015)	Medical device software – Software life cycle processes
Usability	IEC 60601-1-6:2010 +A1:2015 +A2:2020	Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability
	EN 62366-1: 2015+AMD1:2020	Medical devices – Part 1: Application of usability engineering to medical devices
Clinical	IEC 81060-2:2013	Non-Invasive sphygmomanometers - Part 2 Clinical investigation of intermittent automated measurement type
Biocompatibility (System)	ISO 10993-1: 2018	Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process
Risk Management	ISO 14971:2019	Medical devices — Application of risk management to medical devices
Quality System	ISO 13485:2016	Medical devices – Quality management systems – Requirements for regulatory purposes
Symbols	ISO 15223-1:2021	Medical Devices – Symbols to be used with information to be supplied by the manufacturer – Part 1: General requirements
Information	ISO 20417:2021	Medical devices - Information to be supplied by the manufacturer