









Declaration of Conformity

 SRN	SunTech Medical, Inc. 507 Airport Boulevard, Suite 117 Morrisville, NC 27560-8200 suntechmed.com USA US-MF-000002189	 SRN	EMERGO EUROPE Prinsessegracht 20 2514 AP The Hague The Netherlands NL-AR-000000116
	Product Name All Purpose Cuff (APC) One Piece Durable Cuff (OPC) APC: 222APC OPC: 222OPC	APC Basic  OPC Basic  	08409351000000000222APCF4 08409351000000000222OPCHA See attachment
Description Intended Purpose	All Purpose and One-Piece Durable Blood Pressure Cuffs The All-Purpose Cuffs and One-Piece Durable Cuffs are intended to be used with a manual or automatic non-invasive sphygmomanometer to determine blood pressure parameters on pediatric and adult patients. These cuffs are not intended to be used on neonates. These cuffs are intended to be used with prescription (clinical grade) blood pressure monitors or with over-the counter (home) blood pressure monitors.		
Classification	Class I, Rule 1	Assessment Procedure	Annex II and Annex III
Notified Body	N/A	Product Marking	
GMDN Code and Term	34978 - Blood pressure cuff, reusable	UMDNS Code and Term	11703- Devices that have an inflatable bladder in an inelastic sleeve (cuff) with a mechanism for inflating and deflating the bladder. These devices are used in conjunction with another device to determine a patient's blood pressure.

This declaration of conformity is issued under the sole responsibility of SunTech Medical Inc. The above system complies with MDR 2017/745 requirements, in accordance with Annex I (General Safety and Performance Requirements), Annex II (Technical Documentation), Annex III (Post-Market Surveillance), and Annex IV (EC Declaration of Conformity) and with WEEE Directive 2012/19/EU, and the European ROHS Directive 2015/863. This declaration is supported by the Quality System approval to ISO 13485 issued by Intertek. All supporting documentation is retained at the premises of the manufacturer.

I, the undersigned, declare on the basis of the above information that the system described above is in compliance with the requirements of MDR Directive 2017/745. This declaration hereby authorizes the CE Mark to be affixed to the above-mentioned product(s).

Reviewed and Approved by:  Date: 10/20/2021
 Randy Evers
 Regulatory Engineer
DocuSigned by:
Randy Evers
Signer Name: Randy Evers
Signing Reason: I approve this document
Signing Time: 10/20/2021 | 10:09:01 AM PDT
C7B4F1E9CA084CBFB500A9ED713EC301

Signed at SunTech Medical, Inc, Morrisville, NC 27560-8200
 Document Expiry Date: 10/20/2022 (maximum of 1 year upon release)

Attachment to Declaration of Conformity

Device variants

REF	Description
98-06XX-XX	On Piece Durable Cuff, various sizes, with 1 or 2 tubes, with various connectors and package sizes. (where -XX indicates any alphanumeric character 0 to 9 or A-Z)
98-059X-XX	
98-008X-XX	
98-009X-XX	
98-01XX-XX	

Standards Applied:

Safety	IEC 80601-2-30:2009 + A1:2013	Medical Electrical Equipment - Part 2-30: Particular Requirements For The Basic Safety And Essential Performance of sphygmomanometers
	EN1060-3: 1997 + A2: 2009	Non-invasive sphygmomanometers-Part 3: Supplementary requirements for electro-mechanical blood
Biocompatibility	EN ISO 10993-1:2018	Biological Evaluation of Medical Devices-Part 1: Evaluation and testing within a risk management process
	EN ISO 10993-5:2009	Biological evaluation of medical devices – Part 5: Tests for in vitro cytotoxicity
	ISO 10993-10:2010	Biological evaluation of medical devices-Part 10: Tests for irritation and skin sensitization
Symbols	ISO 15223-1:2016	Medical Devices – Symbols to Be Used With Medical Device Labels, L Labelling, and Information To Be Supplied - Part 1: General Requirements
Quality System	EN ISO 13485:2016	Medical Devices – Quality Management Systems – Requirements for Regulatory Purposes
Risk Management	EN ISO 14971:2012	Medical Devices Risk Analysis