

Attachment to Declaration of Conformity

Device variants

REF	Description
98-0400-XX	Disposable Cuff, various sizes, with 1 or 2 tubes, with various connectors and package sizes. (where -XX indicates any alphanumeric character 0 to 9 or A-Z)
98-0700-XX	Single Patient Use Kits (where -XX indicates any alphanumeric character 0 to 9 or A-Z)

Standards Applied:

Safety	IEC 80601-2-30:2009 + A1:2013	Medical Electrical Equipment - Part 2-30: Particular Requirements For The Basic Safety And Essential Performance of sphygmomanometers
	EN1060-3: 1997 + A2: 2009	Non-invasive sphygmomanometers-Part 3: Supplementary requirements for electro-mechanical blood
Biocompatibility	EN ISO 10993-1:2018	Biological Evaluation of Medical Devices-Part 1: Evaluation and testing within a risk management process
	EN ISO 10993-5:2009	Biological evaluation of medical devices – Part 5: Tests for in vitro cytotoxicity
	ISO 10993-10:2010	Biological evaluation of medical devices-Part 10: Tests for irritation and skin sensitization
Symbols	ISO 15223-1:2016	Medical Devices – Symbols to Be Used With Medical Device Labels, Labelling, and Information To Be Supplied - Part 1: General Requirements
Quality System	EN ISO 13485:2016	Medical Devices – Quality Management Systems – Requirements for Regulatory Purposes
Risk Management	EN ISO 14971:2012	Medical Devices Risk Analysis